

Minutes of the pre-bid meeting for tender # 3421000008

Supply of medical equipment

On Thursday, 04/02/2021, pre-bid meeting with the potential suppliers was conducted at UNRWA Gaza Procurement & Logistics Department (PLD) meeting room @ 10:00 am.

The meeting was conducted with the presence of the followings:

- 1- Fawzi El Madhoun (Procurement)
- 2- Manal Bader (Procurement)
- 3- Moh'd Naser (Health)
- 4- Ahmed Bashir (MOH)
- 5- Majdi Abu Seedo (MOH)
- 6- Potential suppliers (11 suppliers)

During the meeting, the below points were discussed:

- 1- All tender related terms and conditions were discussed.
- 2- Suppliers were asked to provide clearly the exact delivery terms for each quoted item.
- 3- The tender closure date has been extended to read **16.02.2021** instead of 09.02.2021
- 4- Suppliers were informed to submit their bids electronically to the TOC e-mail which designated in the tender documents and on time.
- 5- The specifications of the required items were discussed one by one and it was agreed on amendments for the following items as follows:

Item # 1: (Syringe pump)

- 1- It should include IV stand

Item # 2: (Flow meter)

- Connector outlet should be OHAO
- Flow rate should be from 0-15 L/M
- Offer should include humidifier bottle

Item # 3 (Ventilator)

- Tidal volume should include age neonatal from 5-2000 ml/m.
- Respiratory rate up to 120 r/m.
- Flow sensor should be multiuse for at least 3 months (life time) and can be auto clavable.

item # 4 (patient monitor)

- Two central station hardware with PC can be locally, but software should be originally.

Item # 5 (patient bed)

- Stainless steel or galvanized steel.
- up/down spring gas system, not hydraulic.
- bed base should be metal.

